



We can only guarantee the performance and safety of our instruments if the appropriate HEINE bulb and battery powers it. If a non-original copy of a HEINE bulb or battery is used, the safety and performance of the HEINE instrument will be compromised and the manufacturer's guarantee will no longer be valid.

Insist on genuine HEINE replacement bulbs, batteries and parts to ensure performance of your HEINE precision instrument.

HEINE batteries

For optimum performance and cost effectiveness

X-02.99.380 NiCad BETA battery has been replaced by X-07.99.381 or X-02.99.382
X-04.99.623 has been cancelled, use X-04.99.624 or X-04.99.641 in Accubox-II / II L units only.

	Instrument	2.5 V	3.5 V
 	BETA LI-ION, Lithium-Ion (LI-ION) 3,5V rechargable battery for: All BETA handles Laryngoscope handles		X-07.99.381
	BETA NiMH, Nickel Metal Hydride (NiMH) 3,5V rechargable battery for: All BETA handles Laryngoscope handles		X-02.99.382
	M3Z, Nickel Metal Hydride (NiMH) 3,5V rechargable battery for: BETA slim handle, Alpha+ handle Small laryngoscope handles		X-02.99.106
	M2Z, Nickel Cadmium (NiCad) 2,5V rechargable battery for: BETA slim handle, Alpha+ handle Small laryngoscope handles	D-01.89.013	
	S2Z, Nickel Metal Hydride (NiMH) 2,5V rechargable battery for: Old S2Z handles	X-01.99.333	
	S3Z, Nickel Metal Hydride (NiMH) 3,5V rechargable battery for: Old S3Z handles		X-02.99.314
	Stand N, Nickel Metal Hydride (NiMH) 2,5V rechargable battery for: Old Nicatron N, Stand N, handles	X-01.99.433	
			6 V
	Nickel Metal Hydride (NiMH) 6V rechargable battery for: Old Accubox-II and Accubox-II L		X-04.99.641
	HC6V, Nickel Cadmium (NiCad) 6V rechargable battery for: Old Accubox-II and Accubox-II L		X-04.99.624
 	Lithium-Ion (LI-ION) 6V rechargable battery for: mPack		X-07.99.675